

US FDA UDI Implementation Update & Lessons Learned

January 27, 2022

Indira Konduri

Deputy Director, Division of Market Intelligence Office of Regulatory Programs Office Product Evaluation & Quality CDRH, US FDA








- September 2013: UDI Final Rule published
- September 2022: All applicable devices should have:
 - -UDI on label &
 - Data in Global Unique
 Device Identification
 Database (GUDID)

>3.2 million records in GUDID




Data Current as of January 10, 2022






Most GUDID Records are Class II



~35% are Associated with Implantables


"Implantable" devices are those assigned FDA Product Codes associated with Implantable Devices, Systems and Accessories



Medical Specialties in GUDID



Data Based on FDA Product Codes Data Current as of Jan 10, 2022



Most Frequent FDA Product Code Terms in GUDID



Data Current as of January 10, 2022





AccessGUDID



- https://accessgudid.nlm.nih.gov/
- Public Access portal for GUDID data
- Managed by National Library of Medicine
- Daily data updates
- Search/Retrieval/Download of device data
- Useful APIs to query device information



GUDID Usage Data for December 2021



Traffic Stats

Users

45,439

■ -5.6%

Sessions

131,777

! -4.3%

Pageviews

976,788

₽ -6.5%

Homepage Pageviews


₽ -7.1%

Avg. Sessions per Day

4,250.87

1 -7.4%

Location



	Country	Sessions -
1.	United States	83.7K
2.	India	23.1K
3.	Canada	2.4K
4.	Philippines	2.4K
5.	China	1.7K
6.	Germany	1.4K
7.	Japan	1.3K
8.	United Kingdom	1.3K
9.	Vietnam	1K
10.	South Korea	828

United States Sessions 83.7K -3.6%	International Sessions 48.1K
United States Pageviews 521.1K	International Pageviews 455.7K





Download & API Usage December 2021

How many HTTP files were downloaded last month?


Total Events

1,466

1-10.8%

Unique Events
1,298
1-5.9%

API Request Analysis (Last month, Log data)





Lessons Learned

UDI Policy



- Strive to Harmonize!! Leverage IMDRF work
- Establish realistic and achievable deadlines for industry & regulatory body
- Make policy & technical decisions based on real world use of UDI data
- Plan for challenging device types: kits, device systems, SaMD etc.
- Consider compliance policy for upkeep and maintenance of device identification data



Device Identification Database

- Strive to Harmonize!! Leverage IMDRF work
- Spend time to establish clear definitions for data elements
- Use standard list of values
 - Consider values specific to each device area, e.g., Device Size
 - Yes/No drop downs vs. check boxes
- Consider how data will be used during system design
 - Begin pilot projects as data starts coming in
- Set up efficient and easy processes to enable data updates
- Establish strong data validation rules

Stakeholder Outreach & Education



- Early & often: listening sessions, webinars, usergroups, FAQs etc.
- Leverage external stakeholders to multiply common messages
- Establish a help-desk for stakeholder inquiries

Adoption



- Initiate UDI-use/value demonstration projects throughout the supply chain and at point of care
 - Document & share demonstrated value & lessons learned
- Support & encourage learning communities and user-groups
- Establish processes to prioritize and take action on noted gaps

Moving Global UDI Efforts Forward





identification

UDI in Regulatory data (annual reports, adverse events, recalls, etc.) repository of device

UDI in Real World Data sources (EHRs, Registries)

Fully leverage UDI for patient safety

US FDA UDI Program Resources

- FDA UDI Website: www.fda.gov/udi
- Public access to GUDID data:
 - AccessGUDID: https://accessgudid.nlm.nih.gov/
 - Open FDA: https://open.fda.gov/apis/device/udi/

- Contacts:
 - Indira Konduri, indira.konduri@fda.hhs.gov
 - Michael Simpson, michael.simpson@fda.hhs.gov



It always seems impossible until it's done

-Nelson Mandela