



**EIECEE**

# Evaluación de la Conformidad

**Steven Margis**

**Director: EIECEE**

Director de UL, Programa de  
Evaluación de la Conformidad

**Utilización de normas voluntarias de  
consenso y evaluación de la conformidad**

17-03-2022


Seminario Web sobre Dispositivos Médicos – Sesión IV



International  
Electrotechnical  
Commission

# Evaluación de la Conformidad

“demostración de que se cumplen los **requisitos especificados** relativos a un producto, proceso, sistema, persona u organismo”



# Requisitos Especificados


Evaluación de la Conformidad

**“necesidad o expectativa que se declara”**


*NOTA: Los requisitos especificados pueden indicarse en documentos normativos tales como **regulaciones**, **normas** y **especificaciones técnicas***

# Evaluación de la Conformidad

“demostración de que se cumplen los requisitos especificados”




Fuente: ISO con respecto a ISO/IEC 17000: 2019





## Sistema IEC de Esquemas de Evaluación de la Conformidad para Equipos y Componentes Electrotécnicos

- **Facilitar el comercio** eliminando la duplicación de pruebas y proporcionando acceso a los mercados
- Adopción y uso nacional de las **normas IEC** (con o sin diferencias)
- **Aceptación mutua** de los certificados EIECEE y sus informes de ensayo relacionados
- **Evaluación por pares** para garantizar la competencia, la coherencia y la confianza mutua






	Argentina		Grecia		Federación Rusa
	Australia		Hungría		Arabia Saudita
	Austria		India		República de Serbia
	Bahréin		Indonesia		Singapur
	Bielorrusia		Israel		Eslovaquia
	Bélgica		Italia		Eslovenia
	Brasil		Japón		Sudáfrica
	Bulgaria		Kenia		España
	Canadá		República de Corea		Suecia
	Chile		Malasia		Suiza
	China		México		Tailandia
	Colombia		Holanda		Turquía
	Croacia		Nueva Zelanda		Ucrania
	República Checa		Nigeria		Emiratos Árabes Unidos
	Dinamarca		Noruega		Reino Unido
	Finlandia		Pakistán		Estados Unidos
	Francia		Polonia		Vietnam
	Alemania		Portugal		








**BATT**  
Batteries



**CABL**  
Cables and cords



**CAP**  
Capacitors as components




**CONT**  
Switches for appliances and automatic controls for electrical household appliances



**E3**  
Electrical energy efficiency



**ELVH**  
Electrical vehicles



**EMC**  
Electromagnetic compatibility



**HOUS**  
Household and similar equipment



**INDA**  
Industrial automation



**INST**  
Installation accessories & connection devices



**ITAV**  
Information technology audio video




**LITE**  
Luminaires



**MEAS**  
Measuring instruments



**MED**  
Electrical equipment for medical use



**MISC**  
Miscellaneous



**OFF**  
IT and office equipment



**POW**  
Low voltage, high power switching equipment



**PROT**  
Installation protective equipment



**PV**  
Photovoltaics



**SAFE**  
Safety transformers and similar equipment




**TOOL**  
Portable tools



**TRON**  
Electronics, entertainment

- 54 Órganos Miembro
- 91 Organismos de Certificación
- 556 Laboratorios de Pruebas
- 1794 Instalaciones de Pruebas para Clientes
- Más de 120.000 certificados (2020)
  - Más de 1.5 millones emitidos
- **MED**
  - 28 Órganos Miembro
  - 51 Organismos de Certificación
  - 148 Laboratorios de Pruebas
  - 275 Normas
  - 2.8% de todos los certificados (3503)



# MED: Economías de los NCB

<i>Bélgica*</i>	Israel	Singapur
Bielorrusia (R*)	<i>Italia*</i>	<i>Eslovaquia*</i>
Canadá	Japón	<i>Eslovenia*</i>
<i>República Checa*</i>	República de Corea	<i>España*</i>
<i>Dinamarca*</i>	<i>Holanda*</i>	<i>Suecia*</i>
<i>Finlandia*</i>	<i>Noruega**</i>	Suiza**
<i>Francia*</i>	Polonia (R*)	Turquía (R*)
<i>Alemania*</i>	Portugal (R*)	Ucrania
<i>Hungría*</i>	Serbia (R)	<i>Reino Unido</i>
		Estados Unidos

Key: Países Miembro de IMDRF (*EU\**, *EEA\*\**)

R: Solo Reconocer

*[no participan Australia, Austria, Brasil, China, Federación Rusa]*



# MED



IEC 60601-1

**INTERNATIONAL  
STANDARD**


**NORME  
INTERNATIONALE**

Medical electrical equipment –  
Part 1: General requirements for basic safety and essential performance

Appareils électromédicaux –  
Partie 1: Exigences générales pour la sécurité de base et les performances  
essentielles

**Ediciones:  
1998 & 2005**





## IECEE OD-2044

Edition 2.3 2019-06

### IECEE OPERATIONAL DOCUMENT


---

IEC System of Conformity Assessment Schemes for Electrotechnical Equipment and Components (IECEE System)


---

Committee of Testing Laboratories (CTL)

Evaluation of Risk Management in medical electrical equipment according to the IEC 60601-1 & ISO/IEC 80601-1 Series of Standards



IECEE OD-2044: 2019-06



## IECEE OD-2055

Edition 2.1 2017-05-17


### IECEE OPERATIONAL DOCUMENT

---

IEC System of Conformity Assessment Schemes for Electrotechnical Equipment and Components (IECEE System)


---

Guideline Document on Medical Electrical Equipment in the CB Scheme according to the IEC 60601 and IEC/ISO 80601 Series of Standards




IECEE OD-2055: 2017-05

CTL DECISION SHEET (DSH)			
Standard(s) (incl. year)	Subclause(s)	Tracking No.	Year
IEC 60601-1: 2005 (ed 3.0)	8.9	2017	2015
<b>Category</b>			
MED			
<b>Subject</b>		<b>Keywords</b>	<b>Developed by</b>
Creepage distance		Creepage distance Working voltage, PCB traces	ETF-3
<b>Approved at</b>			2016 CTL Primary Meeting
<b>Question</b>			
How should the creepage distance be measured in a scenario where the point with the highest working voltage is behind a point with a lower working voltage?			
See the case study below for details:			
Creepage distance requirement on a Class II switch mode power supply unit designed for EMCPP			
A working voltage of 240 Vrms was measured between the secondary side and the nearest printed track on the primary side.			
However, behind the nearest printed track on the primary side there is a point with a working voltage of 340 Vrms.			
The creepage distance between secondary and the nearest 240 V point is 8 mm, which meets the requirement for 240 V.			
The creepage distance between the 340 V point and the 240 V point is 1 mm.			
<b>Question:</b>			
How shall the creepage distance between primary and secondary be measured in such a case?			




Grey dot = solder  
 Yellow dot = solder in line with other solder  
 Red dot = solder in line with location on track

TESTING AND MEASURING EQUIPMENT/ALLOWED SUBCONTRACTING IEC 60601-1:2005 + Am.1:2012 Medical electrical equipment - Part 1: General requirements for basic safety and essential performance			
Clause	Measurement/testing	Testing / measuring equipment / material needed	Subcontracting
4.11	Power input	Suitable devices for the voltage, current, power and frequency Supply: 1 phase and 3 phase variacs	R
5.3	Ambient temperature, humidity, atmospheric pressure	Suitable devices for recording ambient temperature, humidity, atmospheric pressure	R
5.7	Humidity preconditioning treatment	Environmental conditions, Climate chamber controlling temperature and humidity	R
5.9.2	Accessible parts	Force gauge (30 N), standard test finger (figure 6), etc.	R

		Ref. Certif. No.
IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST CERTIFICATES FOR ELECTRICAL EQUIPMENT (IECEE) CB SCHEME		
<b>CB TEST CERTIFICATE</b>		
Product	<input type="checkbox"/> Additional information on page 2	
Name and address of the applicant		
Name and address of the manufacturer		
Name and address of the factory		
<small>Note: When more than one factory, please report on page 2</small>		
Ratings and principal characteristics		
Trademark / Brand (if any)		
Customer's Testing Facility (CTF) Stage used		
Model / Type Ref.		
Additional information (if necessary may also be reported on page 2)		
<small>A sample of the product was tested and found to be in conformity with</small> <small>As shown in the Test Report Ref. No. which forms part of this Certificate</small>	<input type="checkbox"/> Additional information on page 2	
This CB Test Certificate is issued by the National Certification Body		
Date:	Signature:	

Test Report issued under the responsibility of:



IEC 60601-1  
Medical electrical equipment

**Part 1: General requirements for basic safety and essential performance**

Report Reference No. ....: [ ]

Date of issue .....: .....

Total number of pages .....: .....

CB Testing Laboratory .....: .....

Address .....: .....

Applicant's name .....: .....

Address .....: .....

Test specification:

Standard .....: IEC 60601-1:2005, COR1:2006, COR2:2007, AMD1:2012 (or IEC 60601-1:2012 reprint)

Test procedure.....: CB Scheme

Non-standard test method.....: .....

Test Report Form No.....: IEC60601\_1N

Test Report Form Originator.....: UL(US)

Master TRF .....: 2019-09-17

Copyright © 2019 Worldwide System for Conformity Testing and Certification of Electrotechnical Equipment and Components (IECEE), Geneva, Switzerland. All rights reserved.

This publication may be reproduced in whole or in part for non-commercial purposes as long as the IECEE is acknowledged as copyright owner and source of the material. IECEE takes no responsibility for and will not assume liability for damages resulting from the reader's interpretation of the reproduced material due to its placement and context.

If this Test Report Form is used by non-IECEE members, the IECEE/IEC logo and the reference to the CB Scheme procedure shall be removed.

**This report is not valid as a CB Test Report unless signed by an approved CB Testing Laboratory and appended to a CB Test Certificate issued by an NCB in accordance with IECEE 02.**

**General disclaimer:**  
The test results presented in this report relate only to the object tested. This report shall not be reproduced, except in full, without the written approval of the Issuing CB testing laboratory. The authenticity of this Test Report and its contents can be verified by contacting the NCB, responsible for this Test Report.

4	<b>GENERAL REQUIREMENTS</b>		
4.1	Requirements of this standard applied in NORMAL USE and reasonably foreseeable misuse		
4.2	RISK MANAGEMENT PROCESS FOR ME EQUIPMENT OR ME SYSTEMS		
4.2.2	General requirement for RISK MANAGEMENT - PROCESS complies with ISO14971 (2007) .....	See Appended RM Results Table 4.2.2.	
4.2.3	Evaluating risk		
4.2.3.1	a) Compliance with the standard reduces residual risk to an acceptable level		
	b) Manufacturer has defined risk acceptability criteria in the RISK MANAGEMENT PLAN .....	RISK MANAGEMENT PLAN Document ____	
	c) When no specific technical requirements provided manufacturer has determined HAZARDS or HAZARDOUS SITUATIONS exists.		
	- HAZARDS OF HAZARDOUS SITUATIONS have been evaluated using the RISK MANAGEMENT PROCESS.		
4.2.3.2	MANUFACTURER has addressed HAZARDS OF HAZARDOUS SITUATIONS not specifically addressed in the IEC 60601-1 series.		
4.3	Performance of clinical functions necessary to achieve INTENDED USE or that could affect the safety of the ME EQUIPMENT or ME SYSTEM were identified during RISK ANALYSIS.	RM File Reference to Essential performance: ____	
	- Performance limits were identified in both NORMAL CONDITION and SINGLE FAULT CONDITION.		
	- Loss or degradation of performance beyond the limits specified by the MANUFACTURER were evaluated		
	- Functions with unacceptable risks are identified as ESSENTIAL PERFORMANCE .....	See Appended Table 4.3	
	- RISK CONTROL measures implemented		
	- Methods used to verify the effectiveness of RISK CONTROL measures implemented		
4.4	EXPECTED SERVICE LIFE stated in RISK MANAGEMENT FILE .....		
4.5	Alternative RISK CONTROL methods utilized:		



# Otras Consideraciones

- Ciberseguridad
  - IEC 62443
- *Seguridad Funcional*
  - IEC 61508
- *Competencias del Personal*



# Un enfoque metódico

... a la Regulación de Dispositivos Médicos

- Participar
- Objeto de la Evaluación de la Conformidad
- Requisitos especificados
- Actividad de la Evaluación de la Conformidad
- Sistema/Esquema de Evaluación de la Conformidad
- Optimización Global







**IMDRF** International Medical  
Device Regulators Forum

- Australia [Therapeutic Goods Administration](#)
- Brasil [National Health Surveillance Agency \(ANVISA\)](#)
- Canadá [Health Canada](#)
- China [China Food and Drug Administration](#)
- Unión Europea [European Commission Directorate-General for Internal Market, Industry, Entrepreneurship and SMEs](#)
- Japón [Pharmaceuticals and Medical Devices Agency](#) y el [Ministry of Health, Labour and Welfare](#)
- Rusia [Russian Ministry of Health](#)
- Singapur [Health Sciences Authority](#)
- Corea del Sur [Ministry of Food and Drug Safety](#)
- Estados Unidos de América [US Food and Drug Administration](#)



**Steven T. Margis**  
**Director, CPO and Accreditations**



UL LLC  
333 Pfingsten Road  
Northbrook, IL  
60062 USA  
T: 847.664.3042  
M: 630.816.4537  
E: Steven.T.Margis@ul.com / W: ul.com



**Distinguished  
Member of Technical Staff  
William Henry Merrill Society**

**Steven Margis**  
**Director: EIECEE**  
Director de UL, Programa de  
Evaluación de la Conformidad

**Utilización de normas voluntarias de  
consenso y evaluación de la conformidad**  
17-03-2022  
Seminario Web sobre Dispositivos Médicos – Sesión IV

